

KEY NOTES

1. M5530, SLIT LAMP, IS FOR NAVY PROJECTS ONLY.
2. MOUNT A5145, HOOK, AT 1372 [4'-6"] AFF. PROVIDE 2-A5145'S FOR NAVY PROJECTS ONLY.
MOUNT 1-A5145 AT 1372 [4'-6"] AFF AND 1-A5145 AT 1524 [5'-0"] AFF.
3. F3200, CLOCK, IS FOR NAVY PROJECTS AND IS OPTIONAL.
4. INCREASE THIS DIMENSION AS NEEDED TO COMPLY WITH ADA IF THE DOOR HAS A CLOSER
OR IF 610 [2'-0"] CLEARANCE IS NOT AVAILABLE AT THE STRIKE SIDE OF THE DOOR.
5. M1800, COMPUTER, F0280, CHAIR, AND E0210, WORKSURFACE, ARE OPTIONAL.
6. COORDINATE LOCATION OF M1620, PATIENT CHART HOLDER, AND ROOM SIGNAGE.
7. P3200, UTR ELECTRONIC LAVATORY, MAY BE USED INSTEAD OF P3100, LAVATORY. FOR THE P3200
PROVIDE EITHER BATTERY POWER OR HARD WIRED CONNECTION.

EYE, VISUAL FIELD/PERIMETRY ROOM (EYVF1)

OPHTHALMOLOGY/OPTOMETRY


FLOOR/EQUIPMENT PLAN 11.15 NSM (120 NSF)

DATE: 07/12/2001

0 1000 2000



1:50 SCALE (MILLIMETERS)



EYE, VISUAL FIELD/PERIMETRY ROOM (EYVF1)

OPHTHALMOLOGY/OPTOMETRY

CEILING PLAN


11.15 NSM (120 NSF)

DATE: 07/12/2001

0 1000 2000



1:50 SCALE (MILLIMETERS)



EAST

KEY NOTES

1. MOUNT A5145, HOOK, AT 1219 [4'-0"] AFF. PROVIDE 2-A5145'S FOR NAVY PROJECTS ONLY.
MOUNT 1-A5145 AT 1219 [4'-0"] AFF AND 1-A5145 AT 1524 [5'-0"] AFF.
2. P3200, UTR ELECTRONIC LAVATORY, MAY BE USED INSTEAD OF P3100, LAVATORY. FOR THE P3200 PROVIDE EITHER BATTERY POWER OR HARD WIRED CONNECTION.

General Notes:


1. For typical equipment mounting heights see "TYPICAL EQUIPMENT MOUNTING HEIGHTS" drawing.
2. MIL-STD-1691 data can only be accessed from the Equipment Plan.

EYE, VISUAL FIELD/PERIMETRY ROOM (EYVF1)

OPHTHALMOLOGY/OPTOMETRY

ELEVATIONS

11.15 NSM (120 NSF)



DATE: 07/12/2001

1:50 SCALE (MILLIMETERS)

Room Code: EYVF1 Room Qty: 1 Room Area: 120.01 Room Description: EYE, VISUAL FIELD/PERIMETRY

JSN	NOMENCLATURE	Qty	UNIT	LOG	Utl	Utl	Utl	Utl	Utl	Utl
			ISSUE	CAT	1	2	3	4	5	6
Equipment per Room:										
A1066	Mirror, Float Glass, With SS Frame, 36X18	1	EA	A
A5075	Dispenser, Soap, Disposable	1	EA	C
A5080	Dispenser, Paper Towel, SS, Surface Mounted	1	EA	A
A5106	Waste Disposal Unit, Sharps w/Glove Dispenser	1	EA	C
A5145	Hook, Garment, Double, SS, Surface Mounted	1	EA	A
E0210	Worksurface, w/Overhead Cab, Wall Mtd, 48" W	1	EA	C	.	A
F0300	Chair, Typist, Swivel	3	EA	C
F2000	Basket, Wastepaper, Round, Metal, 18 H x 16 Dia.	1	EA	C
F3200	Clock, Battery, 12" Diameter	1	EA	C
M1620	Holder, Chart, Patient, Wall or Door Mounted	1	EA	A
M1800	Computer, Microprocessing, w/CRT Monitor	1	EA	C	.	A
M5025	Table, Instrument, Ophthalmic, Adjustable	2	EA	C	.	A
M5030	Stool, Surgeon, Revolving	2	EA	C
M5530	Lamp, Slit, w/Applanation Tonometer	1	EA	C	.	A
M5550	Perimeter	1	EA	C	.	A
M5555	Perimeter, Automated	1	EA	C	.	A
M5700	Frame Set, Lens, Trial	1	EA	C
P3100	Lavatory, Vitreous China, Slab Type	1	EA	A	D
=====										
GRAND TOTAL		22								