

Certificate of Conformance

Model CEL-480.C1

Serial No

Firmware Version

The instrument identified above has been manufactured, inspected and tested, in accordance with company drawings and specifications, and conforms to the standard(s) indicated below: -

IEC-651: 1979 Type 1

IEC-804: 1985 Type 1

ANSI S1.4: 1983 Type S(1)

IEC-1260: 1995 Class 2

IEC-1252: 1983

IEC-942: 1988 Class 1L

Manufacturers Specification

<input type="checkbox"/>
<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>

Type 1I

Type 2

Type S(2)

Class 2L

<input checked="" type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>

Type 2I

☐

Type 2I

☐

ANSI S1.11: 1986 Type 2

ANSI S1.25: 1992 Class 2

ANSI S1.40: 1984

<input checked="" type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>

Company test equipment and acoustic working standards, used for conformance testing, are subject to periodic calibration, traceable to UK national standards, in accordance with the company's ISO 9001 Quality System.

Signature

Date

14 OCT 2002

CEL instrumentation is manufactured by CASELLA CEL LIMITED.

Regent House, Wolseley Road, Kempston, Bedford, MK42 7JY U.K.

Tel: (44)01234 844 100 FAX (44)01234 841 490 e-mail info@casellagroup.com

17 Old Nashua Road, #15, Amherst, NH 03031, USA

Tel: (603)-672-0031 Toll Free: (800)-366-2966 FAX: (603)-672-8053

e-mail: Sales@CasellaUSA.com



