

Limited Equipment Warranty

Getinge/Castle warrants to the purchaser as limited below that each item of equipment manufactured by and purchased from Getinge/Castle which is installed, operated and maintained in accordance with Getinge/Castle's stated recommendations will be free from defects in material and workmanship, and that each piece of equipment will be in good working order and will conform to Getinge/Castle's published specifications which are available upon request.

1.0 WARRANTY START DATE: The warranty start date will be the earlier of the date of installation or 60 days after shipment. If installation is expected to occur more than 60 days after shipment, the warranty start date can be extended until the date of installation for up to one year from the date of shipment, provided such extension is specifically approved in writing by Getinge/Castle.

2.0 WARRANTY PERIOD: Getinge/Castle equipment is warranted for a period of twelve (12) months from the start of the warranty period, unless otherwise specified below. All defective parts and coincidental labor are provided for under warranty. Warranty service will be provided during normal working hours, 8:00 a.m. to 5:00 p.m., Monday through Friday, excluding national holidays.

2.1 Sterilizer Pressure Vessels - Sterilizer pressure vessels are warranted to the original owner against structural failures as follows:

- 2.1.1 New Sterilizers - Fifteen (15) years
- 2.1.2 Remanufactured Sterilizers - Five (5) years.
- 2.1.3 For either warranty, the vessel shall be repaired or replaced. Replacement vessel exclude on site labor.

2.2 Portable Equipment - Equipment weighing less than one hundred (100) pounds and not requiring tools to disconnect from building services is covered by Getinge/Castle's Customer Ship-In Repair warranty service during the twelve (12) month warranty period. This service requires the Customer to deliver the equipment needing repair to a Getinge/Castle factory service center, prepaid. Following any necessary repairs, Getinge/Castle will ship the equipment back to the Customer's location, FOB Getinge/Castle factory service center.

2.3 Field Modernization Kits - Components supplied by Getinge/Castle as part of Field Modernization Kits are warranted for twelve (12) months, provided they are installed by a Getinge/Castle factory representative. Getinge/Castle does not warrant Components supplied by Getinge/Castle as part of Field Modernization Kits that are not installed by a Getinge/Castle factory representative. Components supplied as part of Field Modernization Kits will be new, remanufactured and will be equivalent to new in performance. Vessels, doors and major customer owned components reused as part of a field installed modernization package are not covered by this warranty.

2.4 Replacement Parts - Parts replaced under warranty will be warranted for the remainder of the original equipment warranty period. The exchange component may be new, remanufactured, reconditioned, repaired or rebuilt, but will be equivalent to new in performance. Components replaced under warranty will become the property of Getinge/Castle, and upon request will be delivered to Getinge/Castle or its authorized representative.

3.0 SERVICES NOT COVERED DURING WARRANTY PERIOD: The following services are not covered under this warranty, however, Getinge/Castle will provide these services, if available, at Getinge/Castle's (i) then generally available hourly service rates and minimum charge for service time, including travel and waiting time, (ii) parts and material prices then generally in effect, and (iii) charges for travel and shipping expenses, all as applicable.

Repair of Equipment damage, replacement of parts or increase in service time caused by:

- a) purchaser's failure to continually provide a suitable environment prescribed by Getinge/Castle, or by improper storage of the Equipment;
- b) purchaser's failure to perform routine or preventive maintenance, as outlined in the Getinge/Castle Equipment owner/operator manual;
- c) neglect or misuse of the Equipment, including use of the Equipment for purposes other than those for which it was designed;
- d) accident or disaster, including fire, water, wind and lightning; vandalism or burglary;
- e) alterations or modifications made to Getinge/Castle's Equipment design;
- f) attachments, including any interconnection to the Equipment of non-Getinge/Castle products or devices not provided under a Getinge/Castle maintenance agreement;
- g) installation, maintenance, or repair of the Equipment performed by other than Getinge/Castle or a service provider authorized by Getinge/Castle;
- h) nicks, dents, scrapes, scratches, or other cosmetic defects, however caused.

4.0 EXCLUSIONS: Expendable Parts (parts that are worn, used or consumed as part of normal operation) and Consumable Parts (parts that must routinely be replaced) are not covered under this warranty. Examples of Expendable Items include: gaskets, trap elements, lubricators, filters, fuses, lamps, bulbs, valve repair kits, bulb holders, etc. Examples of Consumable Items include: printer paper rolls, recording charts, ink, ribbons, detergents, sterilants, biological indicators, chemical indicators, etc.

The warranties under this Agreement are in lieu of any conflicting statement of limited warranty included with an Equipment shipment. Getinge/Castle does not warrant that the operation of the Equipment or programming will be uninterrupted or error-free, or that all programming errors will be corrected.

Damage to the interior of sterilizer chambers or door (s) caused by the use of bleach, acids, or acid vapors, will void the chamber warranty.

THE FOREGOING WARRANTIES ARE IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.


5.0 LIMITATION OF REMEDIES: The extent of the liability of Getinge/Castle for breach of warranty is limited to the repair or correction of defects, the replacement (with a similar item free from the defect in question) of any Equipment which is defective, or the issuance of a credit not to exceed the amount of the original purchase price of the specified product or service which gives rise to the claim, at the option of Getinge/Castle. Such repair, replacement or credit shall be the purchaser's exclusive remedy for breach of warranty.

6.0 LIMITATION OF LIABILITY: Getinge/Castle's liability for damages to the purchaser for any cause whatsoever, and regardless of the form of action, whether in contract or in tort including negligence, shall be limited to the purchase price stated in the applicable contract for the specific Equipment that caused the damages or that are the subject matter of, or are directly related to, the cause of action. The foregoing limitation of liability will not apply to claims of personal injury caused by Getinge/Castle's negligence. In no event, whether as a result of breach of contract, warranty, tort (including negligence) or otherwise, shall Getinge/Castle or its suppliers be liable for any consequential or incidental damages including, but not limited to, loss of profits or revenues, loss of use of any products or any associated equipment, damage to products or equipment, cost of capital, cost of substitute products, facilities, service or replacement service, downtime costs, or claims of the purchaser's own customers for such damages.

7.0 GENERAL: The warranty provided herein and the obligations of Getinge/Castle thereunder shall not be extended, altered or varied except by a written instrument signed by Getinge/Castle and the purchaser.

Contractor Information

- Installed by:
ISEC, Inc.
10601 Calle Lee, Suite 190
Los Alamitos, CA 90720



- Service Provided by Getinge/Castle, Inc.
Contact Customer Service for help at:

800-950-9912

Roland Smith
(as of 7-12-02) is Service Representative for
El Segundo Area




ALM
Castle
MAQUET



RE: Close out for job

1.3.2 Data Package 2

- 
- A hand-drawn black arrow pointing from the left margin towards item 'd' in the list.
- a. Safety precautions: Operators Manual, page 1.
 - b. Normal Operation: Operators Manual, page 5.
 - c. Environmental Conditions: Operators Manual, Appendix A and C.
 - d. Lubrication Data: N/A
 - e. Preventative Maintenance Plan and Schedule: N/A Preventative maintenance actions provided in Service Manual, section 4.
 - f. Maintenance and repair procedures: Service Manual, Sections 5 and 6.
 - g. Removal and replacement instructions: Service Manual, Section 6.
 - h. Spare parts and supply list: Recommended spare parts, Operators Manual, page 9.
 - i. Parts identification: Parts Manual.
 - j. Warranty information: See attached GCI warranty policy.
 - k. Contractor information: N/A

2400 SERIES FIELD HOSPITAL SURGICAL LIGHTS**4. Appendix****APPENDIX B****PERFORMANCE CHECKS**

- Turn ON Switch at Intensity Control.
- Turn Intensity Control Knob Clockwise.
- Turn ON Switch at each Lighthouse. The lights should come on.
- Vary Intensity to see if Illumination varies correctly.
- If all the above are positive, the lights are operational.

APPENDIX C**STORAGE INSTRUCTIONS****COMMERCIAL SHIPPING — IN CORRUGATED CARTON**

- | | | |
|-------------------|---|---|
| • INDOOR STORAGE | - | In Commercial corrugated carton. |
| • OUTDOOR STORAGE | - | None. DO NOT STORE OUTDOORS. |
| • HEIGHT | - | Six (6) cartons high. |
| • HUMIDITY | - | 5 to 95% Relative Humidity. |
| • TEMPERATURE | - | 50 to 100 Degrees F. |
| • FACILITY | - | No special facility. |
| • INSPECTION | - | Periodically perform a visual inspection check.
Suggest once a year. |

MILITARY SHIPPING — IN MILITARY BOX

- | | | |
|-------------------|---|---|
| • INDOOR STORAGE | - | In military box. |
| • OUTDOOR STORAGE | - | In military box and sealed containers. |
| • HEIGHT | - | Four (4) boxes high. |
| • HUMIDITY | - | No restriction. |
| • TEMPERATURE | - | 20 to 110 Degrees F. |
| • FACILITY | - | No special facility. |
| • INSPECTION | - | Periodically perform a visual inspection check.
Suggest once a year. |